



Product Service

Attestation of Compliance

No. E8N 12 11 55975 030

Holder of Certificate: **SENTECH KOREA CORP.**

#63-23 Sinchon-dong
Paju-Si, Gyeonggi-do 413-832
REPUBLIC OF KOREA

Name of Object: **Electrical equ. for measurement, control and laboratory use (Breathalyzer)**

Model(s): **ALP-1**

Description of Object: Rated input voltage: 1.5 V Battery * 2 EA
EMC Classification: Class B equipment according to EN 61326-2-3

Tested according to: EN 61326-2-3:2006

This Attestation of Compliance is issued according to the Directive 2004/108/EC relating to electromagnetic compatibility on a voluntary basis. It confirms that the listed apparatus complies with all essential requirements of the EMC directive and applies only to the sample and its technical documentation submitted to TÜV SÜD Product Service GmbH for testing and certification. See also notes overleaf.

Test report no.: CPSC0140682



Date, 2012-11-20 (Kang-Min Rhee)

CE After preparation of the necessary technical documentation as well as the EC conformity declaration the required CE marking can be affixed on the product. Other relevant directives have to be observed.