



Product Service

Attestation of Compliance

No. E8N 12 08 62545 013

Holder of Certificate: DA TECH CO., Ltd
 391-17, Cheongcheon-dong
 Bupyeong-gu
 Incheon 403-858
 REPUBLIC OF KOREA

Name of Object: Electrical equ. for measurement, control and laboratory use
 (Breathalyzer)

Model(s): DA-8000

Description of Object: Rated input voltage: 3 V d.c. (1.5 V Battery * 2EA)
 EMC classification: Class B equipment
 according to EN 61326-1

Tested according to: EN 61326-1:2006
 EN 61326-2-3:2006

This Attestation of Compliance is issued according to the Directive 2004/108/EC relating to electromagnetic compatibility on a voluntary basis. It confirms that the listed apparatus complies with all essential requirements of the EMC directive and applies only to the sample and its technical documentation submitted to TÜV SÜD Product Service GmbH for testing and certification. See also notes overleaf.

Test report no.: ESTCE1207-003



Date, 2012-08-08 (Kang-Min F.hee)

CE After preparation of the necessary technical documentation as well as the EC conformity declaration the required CE marking can be affixed on the product. Other relevant directives have to be observed.